Template Provided By:

Cancer Imaging Program

National Cancer Institute

Reagent/Supply Specification Sheet

(Refer to GEN-M001)

Reagent/Supply: **Sterile Vented Filter, 0.2 µm**

Storage Conditions: Room Temperature *(if applicable)*

Internal Tracking Code: SS-SVF

MSDS On File? Y N N/A *(circle one)*

Suppliers:

Vendor	Address	Telephone Number	MSDS Available	Vendor Designation (If Needed)
Millipore	**290 Concord Road Billerica, MA 01821**	1-800-645-5476	N/A	A

Ordering Information:

Vendor	Catalog Number	Description	Size/Units	Bulk Quantities	Certificate of Quality Available
Millipore	SLGS V25 5F	Millex-GS 0.22 µm Mixed Cellulose Esters 25 mm with Vented Inlet Ethylene Oxyde Sterilized	50 Filters/Box	N/A	Yes

Internal Specifications: For all filters, the certificate of quality will be compared with the listed internal specifications to ensure that the manufacturers are still testing the filters consistently and utilizing the same standards to determine quality, purity, and uniformity between the filters.

Parameter	Test Methodology	Specification
Bacterial Endotoxins	**< 20 EU/ Device as determined by LAL Filter Integrity**	**Passes Test**
Housing Integrity	**No leakage at 45 psi for 60 seconds**	**Passes Test**
Downstream Particle Count	**< 100 particles ≥ 10 µm**	**Passes Test**

Regulatory Affairs Approval: _____ **Date:** _____

Signature

Cancer Imaging Program, *Confidential--Use and restriction of this document is set forth in Transfer Agreement executed with the National Cancer Institute (NCI).*