Template Provided By:

Cancer Imaging Program

National Cancer Institute

Reagent/Supply Shipping Receipt & Verification

(Refer to GEN-Q003)

Reagent/Supply: Agilent Zorbax SB-Phenyl HPLC column **Expiration Date: _____**

Assigned Internal Tracking Number (ITN): SS-ZOR-L_____

Manufacturer's Assigned Lot Number: _____

Column Performance Report Received? Y N N/A

Inspection of Shipment:

Date of Package Arrival	Visual Inspection Prior to Opening	Visual Inspection Upon Opening	Shipment Acceptable or Returned and Reason Why

Inspected By: _____ Date: _____
Signature

Date Quarantined: _____

Manufacturer's Column Performance Report attached? Y N N/A

For Columns verify the Column Performance Report with the Internal Specifications for Release for Manufacturing

Parameter	Test Methodology	Specification	C of A Verification	Comments
theoretical plates	HPLC of napthalene	>16000 theoretical plates	**Y N N/A**	
selectivity	HPLC of napthalene	1.25-1.35	**Y N N/A**	
USP tailing factor	HPLC of napthalene	< 1.25	**Y N N/A**	

Verified By: _____ Date: _____
Signature

Released By: _____ Date: _____
Signature

Cancer Imaging Program, *Confidential--Use and restriction of this document is set forth in Transfer Agreement executed with the National Cancer Institute (NCI).*