Template Provided By:

Cancer Imaging Program

National Cancer Institute

Reagent/Supply Specification Sheet

(Refer to GEN-M001)

Reagent/Supply: **Econosphere HPLC Column,** *(if applicable)*

Storage Conditions: **Room Temperature**

Internal Tracking Code: SS-ECO

MSDS On File? Y N N/A *(circle one)*

Suppliers:

Vendor	Address	Telephone Number	MSDS Available	Vendor Designation (If Needed)
Alltech Associates, Inc.	**2051 Waukegan Road Deerfield, IL 60015-1899 Fax (847) 948-1078 www.alltechWEB.com**	**(847) 948-8600**	N/A	A

Ordering Information:

Vendor	Catalog Number	Description	Size/Units	Bulk Quantities	Certificate of Quality Available
Alltech Associates, Inc.	**28085**	**Econosphere C18 10 µM HPLC column 10 X 250 mm**	Each	N/A	Yes

Internal Specifications: For all non-USP reagents, the certificate of quality will be compared with the listed internal specifications to ensure that the manufacturers are still testing the product consistently and utilizing the same standards to determine quality, purity, and identity of the supplies.

Parameter	Test Methodology	Specification
Theoretical Plates and Symmetry	HPLC of uracil	> 40000 theoretical plates and symmetry > 0.9
Theoretical Plates and Symmetry	HPLC of phenol	> 40000 theoretical plates and symmetry > 0.9
Theoretical Plates and Symmetry	HPLC of toluene	> 40000 theoretical plates and symmetry > 0.9

Regulatory Affairs Approval: _____ **Date:** _____

Signature

Cancer Imaging Program, *Confidential--Use and restriction of this document is set forth in Transfer Agreement executed with the National Cancer Institute (NCI).*