Template Provided By:

Cancer Imaging Program

National Cancer Institute

Reagent/Supply Specification Sheet

(Refer to GEN-M001)

Reagent/Supply: **^{18}F Separation Cartridge**

Storage Conditions: Room Temperature *(if applicable)*

Internal Tracking Code: SS-FSC

MSDS On File? Y N N/A *(circle one)*

Suppliers:

Vendor	Address	Telephone Number	MSDS Available	Vendor Designation (If Needed)
Macherey-Nagel GmbH & Co. KG	**PO Box 10 13 52 D-52313 Dueren, Germany**	02421-969-0 (Fax) 02421-969-199	N/A	A
GE Healthcare	**P.O. Box 414 Milwaukee, Wisconsin 53201**	1-262-548-5197	N/A	B

Ordering Information:

Vendor	Catalog Number	Description	Size/Units	Bulk Quantities	Certificate of Quality Available
Macherey-Nagel	**731 876**	**Anion Exchange Cartridge (30-PS-HCO_3)**	1 Cartridge	N/A	Yes
GE Healthcare	**731 876**	**Anion Exchange Cartridge (30-PS-HCO_3)**	1 Cartridge	N/A	Yes

NCI Internal Specifications: For all supplies, the certificate of quality will be compared with the listed internal specifications to ensure that the manufacturers are still testing the cartridge consistently and utilizing the same standards to determine quality, purity, and uniformity of the cartridges.

Parameter	Test Methodology	Specification
Microbial Bioburden (aqueous eluate)	Aerobe Bacteria	< 10 cfu*/mL eluate
	Yeast and Molds	< 10 cfu*/mL eluate

Regulatory Affairs Approval: _____ **Date:** _____

Signature

Cancer Imaging Program, *Confidential--Use and restriction of this document is set forth in Transfer Agreement executed with the National Cancer Institute (NCI)*