Template Provided By:

Cancer Imaging Program

National Cancer Institute

Reagent/Supply Specification Sheet

(Refer to GEN-M001)

Reagent/Supply: **Ethanol, USP Absolute, 200 Proof** **Storage Conditions:** Rm Temp., Flammables
(if applicable)

Internal Tracking Code: SS-ETH **MSDS On File? Y N N/A**
(circle one)

Suppliers:

Vendor	Address	Telephone Number	MSDS Available	Vendor Designation (If Needed)
Sigma-Aldrich	**P.O. Box 2060 Milwaukee, WI 53201**	1-800-558-9160 (Fax) 1-800-962-9591	Yes	A
AAPER Alcohol and Chemical Co.	**P.O. Box 339 Shelbyville, Kentucky 40066**	1-800-456-1017 (Fax) 1-502-633-0685	Yes (E200)	B
Any Vendor of Ethanol, USP Absolute, 200 Proof; Use vendors who dehydrate ethanol with molecular sieves only. (Do not use vendors who dehydrate ethanol with azeotropic distillation using benzene as the co-solvent.)				

Ordering Information:

Vendor	Catalog Number	Description	Size/Units	Bulk Quantities	Certificate of Analysis Available
Sigma-Aldrich	**49,354-6**	***Ethyl Alcohol, 200 Proof**	1 pint	N/A	Yes
AAPER	**USP 200 Proof**	***Ethyl Alcohol, USP Absolute 200 Proof**	1 pint	N/A	Yes

NCI Internal Specifications: For all non-USP reagents, the certificate of analysis will be compared with the listed internal specifications to ensure that the manufacturers are still testing the product consistently and utilizing the same standards to determine quality, purity, and identity of the reagents.

***This is an USP Reagent (which is dehydrated using molecular sieves); no further Verification is necessary.**

Regulatory Affairs Approval: _____ **Date:** _____
Signature

Cancer Imaging Program, *Confidential--Use and restriction of this document is set forth in Transfer Agreement executed with the National Cancer Institute (NCI)*