Template Provided By:

Cancer Imaging Program

National Cancer Institute

Reagent/Supply Specification Sheet

(Refer to GEN-M001)

Reagent/Supply: **PTS Endotoxin Cartridge (portable test system)** **Storage Conditions:** **2 to 8°C**
(if applicable)

Internal Tracking Code: SS-PTS **MSDS On File? Y N N/A**
(circle one)

Suppliers:

Vendor	Address	Telephone Number	MSDS Available	Vendor Designation (If Needed)
Charles River Endosafe	**1023 Wappoo Road, Suite 43-B Charleston, SC 29407**	1-800-762-7016 Fax (843) 766-7576	N/A	A

Ordering Information:

Vendor	Catalog Number	Description	Size/Units	Bulk Quantities	Certificate of Analysis Available
Charles River	**PTS2001**	**0.01 EU/mL Sensitivity**	10 per pack	N/A	Yes
Charles River	**PTS2005**	**0.05 EU/mL Sensitivity**	10 per pack	N/A	Yes
Charles River	**PTS201**	**0.1 EU/mL Sensitivity**	10 per pack	N/A	Yes

Internal Specifications: For all non-USP reagents, the certificate of analysis will be compared with the listed internal specifications to ensure that the manufacturers are still testing the product consistently and utilizing the same standards to determine quality, purity, and identity of the reagents.

Check for receipt of the Certificate of Compliance/Analysis to ensure that the Portable Test System (PTS) Cartridge meets the manufacturer's written requirements for an archived curve, negative controls and positive product control results?.

Regulatory Affairs Approval: _____ **Date:** _____
Signature

Cancer Imaging Program, *Confidential--Use and restriction of this document is set forth in Transfer Agreement executed with the National Cancer Institute (NCI).*