Template

"Insert Your Institution Information Here"

Not Responsible for Investigator's Use

Reagent/Supply Specification Sheet

(Refer to Form-M001)

Reagent/Supply: **Acrodisc Syringe Filter, 0.2 µm**

Storage Conditions: Room Temperature *(if applicable)*

Internal Tracking Code: SS-SPF

MSDS On File? Y N N/A *(circle one)*

Suppliers:

Vendor	Address	Telephone Number	MSDS Available	Vendor Designation (If Needed)
Pall Corporation	**25 Harbor Park Drive Port Washington, NY 11050 USA 290 Concord Road Billerica, MA 01821**	**1-800-289-7255**	N/A	A
VWR International	**1310 Goshen Parkway West Chester, PA 19380**	1-800-932-5000	N/A	C

Ordering Information:

Vendor	Catalog Number	Description	Size/Units	Bulk Quantities	Certificate of Quality Available
Pall Corporation	4612	Acrodisc Syringe Filters with Supor Membrane, Sterile	50 filters/case	N/A	Yes
VWR	28143-310	Acrodisc® Syringe Filters with Supor® Membrane, Pall® Life Sciences	50 filters/case	N/A	Yes

[Insert Your Institution Name Here] Internal Specifications: For all filters, the certificate of quality will be compared with the listed internal specifications to ensure that the manufacturers are still testing the filters consistently and utilizing the same standards to determine quality, purity, and uniformity between the filters.

Parameter	Test Methodology	Specification
Bacterial Endotoxins	**< 0.25 EU/ ml as determined by LAL Filter Integrity**	**Passes Test**
Housing Integrity	**No leakage at 45 psi for 60 seconds**	**Passes Test**
Downstream Particle Count	**< 100 particles ≥ 10 µm**	**Passes Test**

Regulatory Affairs Approval: _____ **Date:** _____ _

Signature