Template

"Insert Your Institution Information Here"

Not Responsible for Investigator's Use

Reagent/Supply Specification Sheet

(Refer to Form-M001)

Reagent/Supply: **Gentistic Acid** **Storage Conditions:** RT, Store

(if applicable

Internal Tracking Code: SS-GentA

MSDS On File? Y N N/A
(circle one)

Suppliers:

Vendor	Address	Telephone Number	MSDS Available	Vendor Designation (If Needed)
Sigma-Aldrich	**P.O. Box 2060** **Milwaukee, WI 53201**	1-800-558-9160 (Fax) 1-800-962-9591	Yes	A

Ordering Information:

Vendor	Catalog Number	Description	Size/Units	Bulk Quantities	Certificate of Analysis Available
Sigma-Aldrich	**149357**	ACS reagent grade $\geq$ 98%	10g, 25g, 100g	N/A	Yes

NCI Internal Specifications: For all non-USP reagents (or EuPh), the certificate of analysis will be compared with the listed internal specifications to ensure that the manufacturers are still testing the product consistently and utilizing the same standards to determine quality, purity, and identity of the reagents.

Parameter	Test Methodology	Specification
Appearance	Visual Inspection	White to yellow powder
Purity Acid concentration	Titration (with NaOH)	97.5%-102.5%
Identity	Infrared Structure	Conforms to structure

Regulatory Affairs Approval: _____ **Date:** _____
Signature

DRAFT