Template

"Insert Your Institution Information Here

Not Responsible for Investigator's Use

Reagent/Supply Specification Sheet

(Refer to Form-M001)

Reagent/Supply: **HPLC Analytical Column** **Storage Conditions:** **Room Temperature** *(if applicable)*

Internal Tracking Code: SS-Super **MSDS On File? Y N N/A** *(circle one)*

Suppliers:

Vendor	Address	Telephone Number	MSDS Available	Vendor Designation (If Needed)
GE Healthcare	800 Centennial Avenue P.O. Box 1327 Piscataway, NJ 08855-1327	Tel: 1-800-526-3593 (select option 2) Fax: 1-877-295-8102	N/A	A
any distributor of this column				B

Ordering Information:

Vendor	Catalog Number	Description	Size/Units	Bulk Quantities	Column Performance Report Available
GE Healthcare	17-5175-01	**Superdex 200 analytical column**	each	N/A	

[Insert Your Institution Name Here] Internal Specifications: For all HPLC columns, values from the column performance report will be compared with the listed internal specifications to ensure that the manufacturers are still testing the product consistently and utilizing the same standards for column separation.

Parameter	Test Methodology	Specification
Exclusion limit		1.3×10^6
Separation range		10000-600000 globular protein
Average particle size		13um
Bed volume		24ml

Regulatory Affairs Approval: _____ **Date:** _____

Signature

GE Healthcare Superdex 200 analytical column **Form SS-Super-**

Supersedes: **None** Version: **V1**

Effective Date: _____

Approved by: